

Statement of Memory Volatility and Instrument Security Procedures

Model: Fluke 1550C and Fluke 1555

Product Name: 155X Series MEGOHMMETER

The Fluke 1550C MegOhmMeter is a high voltage insulation tester.
The Fluke 1555 MegOhmMeter is a high voltage insulation tester.

Memory Description:







The 1550C has one memory device:
U9 MSP430F249 60KB+256B Flash Memory, 2KB RAM

The 1555 has one memory device:
U9 MSP430F249 60KB+256B Flash Memory, 2KB RAM

Memory Cleaning Instructions:

Delete Test Results

To delete all saved test results:

1. Push  to call the Function menu.
2. Push  or  to select the menu item **DELETE RESULT**.
3. Push  to call the menu item.
4. Push . **REALLY DEL?** shows on the display.
5. Push  to confirm the deletion.

Note

The Delete function deletes all stored test results. Individual test locations cannot be deleted but are overwritten.