

# Manual Supplement

Manual Title: 279 FC Users  
Print Date: March 2016  
Revision/Date: 1, 5/18

Supplement Issue: **2**  
Issue Date: 4/22  
Page Count: 1

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1, 697

On page 44, replace the **Temperature** section, under **Infrared Camera**:

### **Infrared Camera**

#### **Temperature**

Temperature Measurement Range ..... -10 °C to +200 °C

Temperature Measurement Accuracy .....  $\pm 5$  °C or  $\pm 5$  %, whichever is greater, at 25 °C (Ambient)

for target temperatures below 20 °C, add 0.2 °C for each degree °C

Temperature Coefficient ..... 0.2 °C or 0.2 %, whichever is greater, for each °C from 25 °C

## Change #2, 672

On page 44, replace the **Emissivity** under **Infrared Camera** with:

### **Infrared Camera**

**Emissivity** ..... 0.97 fixed