

# Manual Supplement

**Manual Title:** ii900/ii910 Users Manual  
**Print Date:** April 2019  
**Revision/Date:** Rev. 2, 5/23

**Supplement Issue:** 1  
**Issue Date:** 6/23  
**Page Count:** 1

---

This supplement contains information necessary to ensure the accuracy of the above document.

## Change #1, 330

On page 5, replace:

2. Put one or two batteries into bay of charger base.

with:

2. Put the battery into bay of charger base.

Remove the note after step 4.

## Change #2, 330

On page 10, in *PDQ Mode™ (ii910)*, after the Note, add:

See *LeakQ™ Mode* for instructions on capture and save.

On page 10, in *MecQ™ Mode (ii910)*, after the 1st paragraph, add:

See *LeakQ™ Mode* for instructions on capture and save.

## Change #3, 330

On page 12, in *Tag Note*, step 3a add:

- Action Notes

On page 12, in *Tag Note*, after step 3b add:

### *Note*

*Only the Type of Leak category affects the leak rate estimation. Gas and Pressure are for documentation purposes only.*

## Change #4, 330

On page 15, in *Operating Conditions*, Set PDQ Mode network operating frequency (ii910), remove step 2 and add:

### *Note*

*The PRPD plot is only generated if 50 Hz or 60 Hz is selected.*